

iQM2 Corrective Action Report

October 2025

Analyzer

Model: GEM® Premier 5000
Area: IFLS
Name: IFLS
S/N: 24077514

Cartridge	Date/Time	Event	Corrective Action	Result
Lot No.: 250625G S/N: 0000000005012 20263 Exp. Date: 11/03/2025	10/05/2025 12:32:05 10/04/2025 07:55:15	Cartridge Removed. Samples Remaining = 145. No. of Solution B Adjustments = 0 Solution A Error for Lac	Sensor Output Adjusted	Corrected
Lot No.: 250502R S/N: 0000000005011 69681 Exp. Date: 10/03/2025	10/03/2025 16:02:19	Cartridge Removed. Samples Remaining = 89. No. of Solution B Adjustments = 32		