

iQM2 Corrective Action Report

August 2019

Analyzer

Model: GEM® Premier 5000
Area: VET CLINIC
Name: GP5000
S/N: 18102177

Cartridge	Date/Time	Event	Corrective Action	Result
Lot No.: 190516A S/N: 0000000005000 10975 Exp. Date: 20.09.2019	24.08.2019 07:41:03	Intraspect Detected during sample measurement for Cl. Operator:	Operator Notified. Sensor Value Discarded.	Discarded