

CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

Révision N° : B SOP N° : TEST-F

Page 1 of 1

CODE	INT3104B	LOT	CE20G16L71
		Expiry Date	2022-06

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE20H31L11	Amotosalen	1	190104
assembly	2	N/A		2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE20G16L71	CAD material		00203C3 / 00310C3

DESCRIPTION: INTERCEPT Processing Set for Plasma

CONFORMITY OF TESTS RESULTS

PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)				
Product	Limits (≤)	Date Passed		Results
Final product	0.125 EU / mL	September 7, 2020		PASS
Amotosalen solution sub-assembly	□ 3 mM → 2.0 EU / mL	1	September 7, 2020	PASS
	OR	2	N/A	N/A
	☑ 6 mM → 0.8 EU / mL	3	N/A	N/A
PCT sub-assembly	0.25 EU / mL	1	September 3, 2020	PASS

Disposition : PASS

✓

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09)	Disposition : PASS 🗹
FINAL PHYSICAL TESTING (TEST-C)	Disposition : PASS ☑
pH ANALYSIS ON CONNECTED KIT (TEST-A)	Disposition : PASS ☑
AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01)	Disposition : PASS

STERILITY		
Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑
	Dis	position : PASS

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

2352 UNITS

Reviewed by Fenwal QA MANAGER or DELEGATE

A MANAGER OF DELEGATE

Date & Signature

Released by Cerus QA MANAGER or DELEGATE Date & Signature

> Romain LELIEVRE Sept 21st ,2020

C. BERTHOMIER

September 18, 2020