



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Page 1 of 1

TO BIONIC MEDIZINTECHNIK G.M.B.H.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO DATE OF ISSUE
1027705 31-JUL-25

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-1000	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE	250623002	3,360	23/06/25	JUN2027
811-1009	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT BSB-NP-CLP 16G CE	250612001	1,680	12/06/25	JUN2027
811-1009	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT BSB-NP-CLP 16G CE	250617002	5,010	17/06/25	JUN2027
811-1009	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT BSB-NP-CLP 16G CE	250624001	3,330	24/06/25	JUN2027
814-0302	TRANSFER BAG 300ML CE	250404001	5,000	04/04/25	APR2028

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : BE25/000123,BE25/000011,BE25/000101,BE25/000104,BE25/000126,BE25/000084

CY CHIA
QA MANAGER



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Certificate of Analysis

PRODUCT CODE : 811-1000

LOT NUMBER : 250623002

DESCRIPTION : CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH	5.0 - 6.0	5.7	PASSED
	CITRIC ACID (%)	95 - 105	99.7	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	99.5	PASSED
	* TOTAL CITRATE (%)	95 - 105	102.3	PASSED
	DEXTROSE (%)	95 - 105	100.6	PASSED
	ADENINE (%)	90 - 110	97.6	PASSED

*** TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE**

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 31-JUL-25

CY CHIA

QA MANAGER